

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130579</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>STEVE DUNN</u>	ID# <u>4106</u> Date <u>11-16-10</u> ✓
A Agency <u>APD</u>	Phone # <u>786-8900</u>
Instrument Location <u>ANCHORAGE JAIL</u>	
B Alco S/N <u>22</u> ✓ Target Value <u>.080</u> ✓ High Pressure <u>1100</u>	
Alco Test Values <u>.081</u> ✓ <u>.082</u> ✓	
1st Alco 2nd Alco	
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>RD</sup> day of DEC, 2010.

[Signature]  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579 <sup>BmB</sup> 12/16/10

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130579

NOVEMBER 16, 2010

OPERATOR'S NAME:

DUNN/S

OPERATOR'S NUMBER: 4106

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

V

O.L. #: V

DEPT/AGENCY: ANCI

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .080

ALCO S/N: 2Z

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.45 in

ALCO TARGET .081 12:59

BLANK TEST .000 13:00

INTERNAL STANDARD VERIFIED 13:00

ALCO TV 30.45 in .081 13:00

BLANK TEST .000 13:01

SUBJECT SAMPLE .000 13:02

BLANK TEST .000 13:03

ALCO TV 30.45 in .082 13:03

BLANK TEST .000 13:04

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130579

NOVEMBER 16, 2010

TIME 12:57

---- DIAGNOSTIC CHECK ----

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 12/11/08

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.45 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGH  
IJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~  
pqrstuvwxyz{|}~  
pqrstuvwxyz{|}~